



Pump anytime, anywhere, around anyone.

Equality

Manual Breast Pump



plus the

Hands-free, Concealable Freemie Breast Pump Collection System



User Manual

IMPORTANT
READ THIS MANUAL PRIOR TO USE

This Freemie Equality Manual Breast Pump and Hands-free, Concealable Collection System is a single user, reusable system that you may use throughout the first year of your baby's life, and beyond. This manual contains important information for the proper sanitization, use and care of your Freemie Equality system. Please visit our website, www.Freemie.com, for customer service and all your spare parts needs.

This is a single user product. Use by more than one person may present a health risk and void the warranty.

READ ALL INSTRUCTIONS

Table of Contents

Important Safety Instructions	
Do's and Don'ts	2
Freemie Collection System	
Parts Included	3
Freemie Equality Manual Pump	
Pump Features	4
Preparing Freemie for First Use	5
Assembling Freemie Cups	6
Positioning within Your Bra	
The Perfect Bra Fit.....	7
The 12 O'clock Position	8
Connecting to Your Equality Pump	9
Pumping with Freemie Equality	
Adjusting Your Pump And Sitting Upright	10
When You Are Interrupted	11
Disconnecting and Transferring Milk	
When You Are Done Pumping	12
Pouring Milk Into A Container	13
Cleaning	13
Caring for Your Freemie System	
Tubing Care	14
Troubleshooting Tips	
Breasts not emptying	15
Loss of Suction	15
Milk in tubing	15
Need to stop or break suction	15
Spare Parts & Customer Service	16

Indications for use:

The Freemie Equality Manual Breast Pump and Hands-free, Concealable Collection System is intended to be used for the purpose of expressing human milk.

pump the freemie way

pump efficiently

Single or double pumping

Freemie can be used for single or double pumping as desired.



pump and relax

Home

Simultaneously pump or passively collect milk while breastfeeding to catch and store every precious drop.

Work

Pump discreetly at your work station without ever undressing.

Travel

Travel with Freemie Equality to express milk on the go whenever a private place can't be found.

pump while bonding

Freemie Equality can be used to pump milk while breastfeeding. Place baby on one breast and Freemie cup on the other.

**Not made with BPA,
DEHP, or natural
rubber latex**



Important Notice

If you have any concerns about your health, or experience any pain while nursing or using any breast pump, consult your health care or lactation professional immediately. Dr. Dao, Dao Health and Freemie User Manual do not give medical or lactation advice, or make diagnoses.



Important Safety Instructions

WARNING:

USE FREEMIE HANDS-FREE, CONCEALABLE COLLECTION SYSTEM ONLY WITH FREEMIE COMPATIBLE PUMPS. USE WITH UNAPPROVED PUMPS MAY CAUSE INJURY TO YOU OR DAMAGE YOUR PUMP. MISUSE MAY VOID WARRANTY.

Do

- Use Freemie Hands-free, Concealable Collection System only with compatible pumps.
- Keep Freemie Hands-free, Concealable Collection System in an upright position while pumping.
- Transfer milk to a proper storage container right after pumping.
- Use when you must express milk discreetly and hands-free underneath normal clothing.
- Use Freemie only as often and only as long as you normally breastfeed your baby. Consult a lactation professional if you have questions about your milk supply.
- Use with a bra or manually hold in place.
- Keep out of reach of infants and children. Freemie Equality has small parts that can be a choking hazards.
- Consult your lactation professional if you have questions about whether Freemie Equality is appropriate for you or to help determine the correct funnel size for you.

Do Not

- Do not bend forward or recline while using Freemie cups, and avoid excessive movement. This could cause leakage, loss of suction, and milk to flow into tubing.
- Do not overfill Freemie cups, as this may cause milk to flow into the tubing. You may hear a faint bubbling sound as the Freemie cup is filling. There is a risk of overfilling if you are hearing impaired or use Freemie Equality in a noisy environment.
- Do not use Freemie Equality if you must watch your milk ejection reflex or manually massage your breasts while pumping to ensure that they are completely empty.
- Do not share. Freemie Equality is a personal use item.
- Do not use Freemie Equality in any situation that may be unsafe should you become drowsy or distracted during milk expression.
- Do not use Freemie Equality for simultaneous breastfeeding and pumping with compromised infants without first consulting your healthcare or lactation professional.
- Do not continue to use Freemie Equality if you experience pain or discomfort in the breast or nipple while pumping.




Freemie Breast Pump Collection System

Two cup assemblies included with each Freemie Breast Pump Collection System.

Additional funnel sizes available at www.Freemie.com.



Parts included:



#FG023

Equality Connection Kit



T Connector
(2)



Pressure Reducer
(1)




Freemie Tubing
(4) 24"




Pump Tubing
(1) 32" (80cm)

Equality Manual Pump

One hand pump included with the Equality Manual Pump set.



Pump Features:



Do not cover up hole when pumping



Preparing Freemie for First Use


Disassemble and clean Freemie prior to first use. Some parts must be boiled initially.



Unsnap cup from funnel




Pull valve base off funnel



Remove valve from valve base

Sanitize Parts

**Boil these parts
for 10 minutes
to sanitize**



- Do not let parts sit on the bottom of the boiling pot. Parts should float freely.
- Remove from water and allow to dry completely.

! Parts must be thoroughly dry before use.

**Do NOT boil
these parts**



- Hand wash valve in hot, soapy water and rinse. Allow parts to cool and dry completely before use.

! Do not boil or steam sterilize tubing, valves, or pressure reducer.




Assembling Freemie Cups

Always wash hands prior to assembling Freemie. All parts must be completely dry, correctly positioned, and secure.

Inspect the device before each use for signs of wear or damage and replace parts if signs of wear are noted.

1

Keep stem aligned with tubing port




2

Push Breast Funnel and Valve Base firmly together



3

Place Valve onto Valve Base



4

Make sure Valve is straight




Correct



Incorrect


5

Gently tug on Valve to ensure it is secure




6

Align ports and snap together




7

Insert Tubing into Valve Base



8

Push to ensure tubing is securely in port





Positioning Within Your Bra


The Perfect Bra Fit




Bra fit is critical to the proper use of this device.

1. Before choosing a bra, single pump with Freemie cup using your hand to hold in place. Consider how much pressure you normally apply against your breasts to pump. This pressure may be different when your breasts are engorged versus when they are empty. To achieve the same pressure and sensation with the Freemie cup in your bra, it may be necessary to adjust your bra or try different bras to get the best fit for you. A sports bra works well for some women.
2. Your bra must keep the Freemie cups in an upright position and not allow the device to slip around, sag downward, or tilt off to the side during use.
3. Place Freemie inside your bra with the Breast Funnel opening centered over the nipple.
4. Make sure you achieve a good seal between Freemie and your skin.

Correct Position




Incorrect Position



Ensure nipple is centered in the Breast Funnel and not bent or pinched off.

The 12 O'clock Position

When connected to a pump, tubing must be in the 12 o'clock position.



Let tubing hang down underneath your shirt.



Tubing may be cut to a desired length in order to conceal it under normal clothing.

✓ Correct


When filling with milk, keep in the upright position



Tubing in 12 o'clock position

✗ Incorrect

Placement at any other position can cause milk to flow into tubing








Connecting to Your Equality Pump



Double Pumping



Single Pumping






Pumping with Freemie




Adjusting your pump and sitting upright

1 Adjust pump to low by turning suction control dial to MIN.

2 Slowly pull handle until you feel a comfortable suction. If suction is not adequate, adjust suction control dial toward MAX until you feel a comfortable balance between suction desired and effort required to pull the pump handle. If suction is too strong, adjust suction control dial towards MIN to reduce suction strength. Pull and release repeatedly to stimulate milk letdown, and continue to express milk as needed. If this repetitive motion results in pain, even if you don't think the pump is the cause of pain, stop pumping and consult your health care professional.



3 Pump with Freemie in an upright position, such as sitting or standing.

✓ Correct	✗ Incorrect	✗ Incorrect
<p>Do sit or stand upright</p> 	<p>Do not bend forward</p> 	<p>Do not recline</p> 


4 Each Freemie holds 8 ounces of milk.

NOTE: When device is full, stop pumping immediately and pull tubing from Freemie while still wearing in bra. (see page 12, When Your Are Done Pumping)

5 If you regularly produce more than 8 ounces at a time, stop pumping when Freemie is full and empty device. (See page 12, When Your Are Done Pumping) Alternatively, if you have a second Freemie set, switch out full Freemie cups with empty cups to minimize pumping interruptions.

When you are interrupted

1. Stop pumping.
2. Disconnect tubing at T-connector or at pump.
(See inset on right)
3. Leave Freemie cups temporarily in your bra, and maintain an upright position.
4. When ready to resume pumping reconnect via tubing, and continue to pump.



Important

- Bending over or reclining during use, or overfilling may cause leaking, loss of suction, or milk to flow into tubing.
- Empty Freemie when full and remove immediately after pumping.
- If milk flows into tubing immediately stop pumping. See tubing care on page 14 for cleaning instructions.
- See Troubleshooting sections on page 15 for additional tips.



Stop pumping


if you experience pain in your breast or nipple, even if you think the pump is not the source of the pain.



Disconnecting and Transferring Milk

When you are done pumping

1. Stop pumping.
2. Leave Freemie in your bra and gently pull the tubing from the device.
3. Carefully remove cups from your bra.




4. Freemie cups may be set flat on a table until ready to transfer to a storage container.




5. When Freemie cups are full, do not lay flat on back unless tubing is disconnected.

✘ Incorrect



✔ Correct




Pouring milk into a container



Practice with water first to achieve successful transfers.

Handle a full Freemie cup with care. **Milk can spill out of the Breast Funnel.**

Removal of tubing creates pour spout for transferring milk into storage containers.



Pour quickly with Breast Funnel side up as pictured.

Fully disassemble and clean after each use.

Cleaning

1. Disassemble and wash all components that come into contact with breast milk (cup, breast funnel, valve, valve base).
2. Handwashing in hot soapy water is recommended. Rinse thoroughly.



Do not steam or boil tubing or valves, as it will cause damage.

3. Treat all parts with special care. Avoid abrasives and excessive heat.
4. Clean as soon as possible after each use.
5. When not in use, store parts out of reach from infants and children.




Dry parts completely before safely storing.



Caring for Your Freemie System

Tubing Care

Inspect tubing for breast milk after each pumping session. Soiled tubing that cannot be cleaned should be replaced. Tubing cleaning options are listed below:

1. If breast milk gets into tubing, rinse thoroughly in hot water. Shake excess water from tubing. Wipe with a clean, soft towel and allow to air dry completely while hanging vertically.
2. Rinse tubing with isopropyl alcohol. Hang tubing vertically. Air dry completely.
Caution: Be certain alcohol has completely evaporated before using.
3. If unable to thoroughly clean tubing, cut off any soiled portion of the tubing, or replace it. (see Ordering Parts, page 16)
4.  **Do not clean tubing in dishwasher, microwave steam bag, or by boiling.**



Troubleshooting Tips

Breast Not Emptying

- Always center Breast Funnel over nipple. If the Breast Funnel is off the center of the nipple you may impede milk release, experience loss of suction, or prevent the complete emptying of the breast. (see page 7)
- Try adjusting the position of Freemie. Comfortably position your body to help create the best seal between Freemie and your skin. Experiment with your body position by sitting up straighter, leaning forward a little, or pulling your shoulders back.
- The Breast Funnels come in the standard 25mm, 28mm, and 32mm funnels. If you need a larger or smaller Breast Funnel, visit Freemie.com to order the right size Freemie for you.
- Each Freemie holds up to 8 ounces of milk. If you pump more than this at one time, you need to empty and clean the device to continue pumping. (see page 13)
- If problems emptying your breasts persist, consult your lactation professional.

Loss of Suction

- Make sure all parts are completely dry before assembling.
- Make sure that parts are correctly assembled and fit together as snugly as possible, and that connections to the pump are secure. (see pages 5 & 6 and 9)
- Make sure tubing is not kinked or pinched. Make sure tubing is pushed securely into its port on the valve base. (see page 6)
- Make sure Breast Funnel is securely pressed against breast creating a seal around nipple, and is properly seated in bra. Check that no bra material or tubing is between the device and your skin. The device should not be tilting upward, downward, or to the side. (see pages 8 & 10)
- Make sure nipple is not bent or pinched inside the Breast Funnel. (see page 7)
- Try a tighter fitting bra to ensure that the bra is able to support and maintain an adequate seal and hold Freemie in an upright position. Consider trying a sports bra. (see page 7)

Milk in Tubing

- Always make sure device is placed in bra with tubing and pour spout at the 12 o'clock position. (see page 8)
- Avoid overfilling, empty when full, and remove immediately after pumping.
- Avoid excessive movement while pumping and only use in an upright position. (see page 10)
- Never place full device on flat surface without first removing tubing. (see page 12)
- If milk gets in tubing, see cleaning instructions. Replace tubing if you are unable to clean completely.
- Review your pump manual for pump cleaning instructions.

Need to Stop or Break Suction

- Suction can be broken by stopping the pump or by inserting a finger between the breast and the breast funnel.
- Stop suction if you experience pain in your breast or nipple even if you think the pump is not the source of the pain.



Spare Parts & Customer Service

Freemie Breast Pump Collection System

Part #	Part Name	Picture
FG024	25mm Breast Funnel	Page 3
FG025	28mm Breast Funnel	Page 3
FG026	32mm Breast Funnel	Page 3
FG027	Natural Shape Cup	Page 3
FG029	Valve Base	Page 3
FG030	Valve	Page 3

Freemie Equality Pump Connection Kit

Part #	Part Name	Picture
FG023	Freemie Equality Pump Connection Kit	Page 3



For help with your Freemie product visit us at
www.freemie.com

Customer Service (916) 339-7388
Monday - Friday, 9:00am - 5:00pm PST

PO31.A
AW040.A
10/22/13